

VistA Imaging Approved DICOM Modality Interfaces

November 2024

VistA Imaging has successfully tested many modalities and interfaced them to VistA using DICOM. The following list of approved DICOM modalities was generated based on the most current information and will change as more information becomes available.

A DICOM modality is considered “Supported” if it passes the VA’s internet validation test suite listed in the [Joint VA/DoD DICOM Modality Conformance Requirements \(Rev 3.0\)](#) document. Some modalities have superscript numbers which refer to the notes at the end of the document.

If the software used on a modality is revised by the manufacturer, the modality will need to be re-validated.

The VA retains the right to change modality testing requirements at any time as needed for VA medical care. These changes could require re-testing of the modalities.

VA facilities are required to complete a quality review for the images produced by the device before it may be placed into clinical use. The results of the quality review must be filed with the National VistA Imaging Project Office.

This list of approved DICOM modalities does not in any way constitute a product endorsement by the VA. It only states whether or not the listed product has passed the VA’s internet-based validation testing. Upgrades to the vendor devices can affect interoperability and may require re-validation. The validation process does not indicate how well the actual product will work operationally, either within the VA or elsewhere.

Modality/Device Types

Approved modalities and devices are grouped into the categories listed below. A list of approved devices for each category can be found on the page listed below.

Modality/Device Types.....	1	Ophthalmic Workstation.....	24
Bone Density	2	Optical Coherence Tomography (Ophthalmology use only).....	24
Cardiology	2	Other	25
CD/DVD Burners	2	PET	25
Computed Radiography	3	PET/CT.....	25
Computed Tomography	5	PET/MR.....	26
Dental Extra-Oral Radiography	8	Q/R – 3D Reconstruction Workstation	26
Dental Intra-Oral Radiography	9	Q/R – Planning Workstation.....	26
Dental VL – Endoscopy.....	10	Q/R – Routers	26
Dental Workstations	11	Q/R – Workstations	27
Digital Subtraction.....	11	Radio Fluoroscopy and X-Ray Angiography ...	28
Direct X-Ray.....	11	Secondary Capture.....	31
Evidence Creator (Encapsulated PDF)	15	Secondary Capture (Injection Dosage Capture)	32
Film Digitizer	15	Surgical Workstation	32
Film Digitizer Software	15	Ultrasound	32
Frame Grabber.....	16	Visual Field Analyzer (Encapsulated PDF).....	37
General Microscopy.....	16	VL – Endoscopy	37
Magnetic Resonance.....	16	VL – Endoscopy/Microscopy	38
Mammography.....	19	VL – Pathology.....	38
Middleware – Interface Engine.....	19	XA – X-Ray Angiographic.....	38
Nuclear Medicine.....	21	XC – External-camera Photography (Encapsulated PDF)	39
Nuclear Medicine - SPECT	22	39
Ophthalmic Photography	22	Notes.....	40
Ophthalmic Ultrasound.....	24		

Bone Density

Supported Bone Density Devices

GE DPX Bravo
GE DPX Duo
GE DPX-MD+
GE DPX-NT
GE Lunar iDXA
GE Prodigy Advance
GE Prodigy Primo
GE Prodigy Pro
GE/Lunar Prodigy
Hologic Discovery
Hologic Horizon
Hologic Explorer
Hologic Oasis (QDR) Workstation
Norland Excell (IlluminatusDXA)
Norland XR-46 (IlluminatusDXA)
Norland XR-600 (IlluminatusDXA)
Norland XR-800 (IlluminatusDXA)

Cardiology

Supported Cardiology Devices

GE Innova IGS 520
GE Innova IGS 530
GE Innova IGS 540
GE Innova IGS 620
GE Innova IGS 630
GE Discovery IGS 730
GE Discovery IGS 740
Philips IntelliSpace Cardiovascular version 2.3
Philips Medical Systems AlluraXper
GE Healthcare MUSE 9 (DICOM Storage)

CD/DVD Burners

Supported CD/DVD Burners

All Pro Solutions, DCMburner, v6.14
Pacsgear Media Writer D100/D200¹
Perennity CD Burner Software V4.5
RadInfo Systems CDPowerPACS²
Sencor DICOM Part 10 Burner
Sorna FilmX
Sorna eXpedo
Sorna Vertex
SST Group, DMC-EZ

Notes

- 1 Uses DICOM software that was approved for Pacsgear and Pacscan
- 2 2 Device now supported by Sorna

Computed Radiography

Supported Computed Radiography Devices

Agfa ADC-51xx
Agfa ADC 70⁴
Agfa ADC Compact⁴
Agfa ADC Solo⁴
Agfa ADC-QS
Agfa ADC-VIPS¹
Agfa NX Workstation for all CR Plate Readers
Agfa PRID⁸
Canon Soltus 500 with CXDI, v. 2.19
Carestream Classic
Carestream CR825
Carestream CR850
Carestream CR950
Carestream CR975
Carestream Elite
Carestream POC 360
Carestream Vita CR
Carestream OnSight 3D Extremity System (Version: 1.1)⁹
Carestream (APELEM) DRX-Excel Plus (Version: 1.12.03)
Carestream DRX-Evolution, v.1.5²³
Carestream DRX-Compass, v. 1.6²⁸
Carestream DRX Excel Plus, v. 1.20
Carestream DRX-EVO v.1.9²³
Carestream DRX-Mobile v.1.9²³
Carestream DRX-Revo v.1.9²³
Carestream DRX-RevoPlus v.1.9²³
Carestream DRX-EVO PLUS v.1.9²³
Carestream DRX-Ascend v.1.10²³
Carestream DRX-EVO v.1.10²³
Carestream DRX-EVO PLUS v.1.10²³
Carestream DRX-Revo v.1.10²³
Carestream DRX-RevoPlus v.1.10²³
Carestream DRX-Compass, v. 1.10²³
Carestream DRX-InRoom v.1.10²³
Carestream DRX-Nano v.1.10²³
Carestream DRX-Mobile v.1.10²³
Carestream DRX- Transportable v.1.1023
Carestream DRX- Transportable Lite v.1.1023
Carestream DRX-ImageView, V. 2.10.x
Carestream DRX-Revolution, v. 2.10.x
Carestream DRX-Evolution v. 2.10.x
Carestream DRX-Evolution Plus, v. 2.10.x
Carestream DRX-Rise, v. 2.10.x
Carestream DRX-In-room, v. 2.10.x
Carestream DRX-Transportable/Lite, v. 2.10.x
Carestream DRX-Compass, v. 2.10.x
Fuji 5000⁴
Fuji 5000R⁴
Fuji 7000⁴
Fuji 7501⁴
Fuji 9000⁴
Fuji 9501⁴

Fuji AC1 Plus⁴
Fuji AC3⁴
Fuji Clearview CSm (Platform: Flash IIP vFLEXUIV8.2)
Fuji Clearview CS1m (Platform: Flash IIP vFLEXUIV8.2)
Fuji Carbon
Fuji Carbon Series (Platform: Flash IIP vFLEXUIV8.2)
Fuji FCR Go
Fuji FCR GO (Platform: Flash IIP vFLEXUIV8.2)
Fuji FCR GO 2 (Platform: Flash IIP vFLEXUIV8.2)
Fuji FCR IIP Console (CR-IR348)
Fuji FCR XG1
Fuji Flash IIP
Fuji IDT741
Fuji QA771
Fuji SmartCR
Fuji XG-2000
Fuji XG-2000 (Platform: Flash IIP vFLEXUIV8.2)
Fuji XG-5000
Fuji XG-5000 (Platform: Flash IIP vFLEXUIV8.2)
GE Centricity¹
GE Healthcare Discovery RF180, v. 0.0.3.1²⁸
iCRCo Captera⁵
J. Morita USA, Device: Veraviewepocs 2D, Software: i-Dixel, Version: 2.4.0.4
Kodak CR2000T
Kodak DirectView CR500
Kodak DirectView CR800
Kodak DirectView CR825
Kodak DirectView CR850
Kodak DirectView CR900
Kodak DirectView CR950
Kodak DirectView CR975
Kodak (Carestream Health) DirectView CR System (v5.4)
Konica Image Pilot
Konica Regius 150
Konica Regius CS-3 (0863)
Konica Xpress (CS-1)
Konica Minolta Healthcare America Inc, ULTRA, 5.1.4.9
Lumisys ACR/2000
Orex ACL2 and ACL4⁴
PAUSCH Medical GmbH Uroview FD (V. 8.0.15.15)
Philips AC3⁴
Philips Compano PCR
Philips PCR Corada
Philips PCR Eleva
Philips, Product: Spectral CT Series, Version: 5.2
Siemens AC3⁴
Siemens Multix FD
Siemens Multix Fusion Max
Siemens Multitom Rax, VF10
Siemens Ysio Max, VF10
THALES, Product: ArtPix Mobile EZ2GO, Version: 5.01.00
Viztek Onyx Rad workstation

Notes

1 No DICOM Modality Worklist capability.

- 4 Data is transferred via an intermediary device. Confirm intermediary device was purchased as well.
- 5 Tested as DICOM Object. Test images were not converted to Targa format.
- 8 Only supplies Modality Worklist.
- 9 Storage commitment not tested.

Computed Tomography

Supported Computed Tomography Devices

Canon Medical Systems, Aquilion Precision (V8.80)
 Canon Medical Systems, Aquilion One (V10.00)
 Canon Medical Systems, Cartesion Prime (V10.10ER003)
 Carestream CS 8100 3D Trophy 6.4
 Carestream CS 9300 Trophy 6.4
 Carestream (APELEM) DRX-Excel Plus (Version: 1.12.03)
 Dental Imaging Technology Corporation, Device: i-CAT FLX V-17, Software : SmartScan STUDIO ,Version : 2.13.0.1242
 Elscint CT Twin
 Elscint OmniPro
 GE Brightspeed (ALL)
 GE CT/I¹
 GE Discovery CT660
 GE Discovery CT750 HD
 GE Discovery 590 (RT)
 GE Gensys
 GE HiSpeed CT/I
 GE HiSpeed RP
 GE Jupiter
 GE Lightspeed (ALL)
 GE Optima 520/540
 GE Revolution
 GE Revolution Apex, v.22a.10²³
 GE Revolution Evo
 GE Revolution HD
 GE Revolution GSI
 GE HealthCare,Product: Revolution Ascend, Software: Revolution Scanner²⁸ , Version: 23MW17.18
 GE HealthCare,Product: Revolution CT, Software: Revolution Scanner²⁸ ,Version: 23MW17.18
 GE HealthCare,Product: Revolution Apex, Software: Revolution Scanner²⁸ ,Version: 23MW17.18
 GE HealthCare,Product: Revolution CT / Apex, Software: Revolution Scanner²⁸ , Version: 23MW24.44
 GE HealthCare,Product: Revolution Ascend, Software: Revolution Scanner²⁸ , Version: 23MW24.44
 GE HealthCare,Product: Revolution Ascend Elite, Software: Revolution Scanner²⁸ , Version: 23MW24.44
 GE HealthCare,Product: Revolution Ascend Plus, Software: Revolution Scanner²⁸ , Version: 23MW24.44
 GE HealthCare,Product: Revolution Ascend Select, Software: Revolution Scanner²⁸ , Version: 23MW24.44
 J. Morita USA, Device: Veraview X800, Software: i-Dixel, Version: 2.4.0.4
 J. Morita USA, Device: Veraviewepocs 3D, Software: i-Dixel, Version: 2.4.0.4
 Hitachi Medical Corporation Scenaria
 Hitachi Healthcare Americas, Scenaria View, v. 0005
 Marconi PQ2000
 Marconi PQ5000
 Marconi PQ6000
 Marconi PQS
 Marconi TomoScan^{1,3}
 Marconi Ultra Z¹⁰
 Marconi/Philips Extended Brilliance
 Marconi/Philips MX8000
 Marconi/Philips MX8000 IDT
 NeuroLogica Corp. BodayTom Portable CT

NeuroLogica Corp. CereTom Portable CT
 NeuViz CT16
 NeuViz NV64
 Philips Brilliance 40 CT
 Philips Brilliance 64 CT
 Philips Brilliance CT
 Philips Brilliance iCT
 Philips Brilliance Ingenuity CT
 Philips Brilliance Portal¹⁷
 Philips Brilliance Workspace¹⁷
 Philips Extended Brilliance Workspace¹⁷
 Philips MxViewEXP
 Philips Medical Systems, Philips IQon, 4.7.7
 Philips Medical Systems, Philips Brilliance Big Bore, 4.8
 Philips Medical Systems, CT Incisive, 5.0
 Philips Medical Systems, Philips Spectral CT 7500, v5.0
 Philips Medical Systems, Device: iCT 256, Version: 4.1.10
 Philips, Product: CT 5300, Software: INCISIVE, Version: 5.1
 Riverain Technologies Inc, Device: ClearRead³⁶, Version: 5.2.0.26
 Siemens Biograph_Horizon (V. VJ30B)
 Siemens Biograph mCT (Version VG70A)
 Siemens Biograph mCT (Version VG80A)³²
 Siemens CT Definition
 Siemens Definition Flash
 Siemens Emotion 6
 Siemens Emotion 16
 Siemens Sensation 10
 Siemens Sensation 16
 Siemens Sensation Cardiac
 Siemens SOMATOM1,3
 Siemens SOMATOM Plus1,3
 Siemens Volume Zoom
 Siemens SOMATOM Confidence (version: VA62A)
 Siemens SOMATOM Definition (version: VB10A)
 Siemens SOMATOM Definition Flash (version: VB10A)
 Siemens SOMATOM Definition AS (version: VB10A)
 Siemens SOMATOM Definition Edge (version: VB10A)
 Siemens SOMATOM Drive (version: VB10A)
 Siemens SOMATOM Edge Plus (Version: syngo CT VB10A)
 Siemens SOMATOM Force (version: VB10A)
 Siemens SOMATOM Definition Edge (version: syngo CT VB20A)
 Siemens SOMATOM Definition Flash (version: syngo CT VB20A)
 Siemens SOMATOM Definition AS (version: syngo CT VB20A)
 Siemens SOMATOM Drive (version: syngo CT VB20A)
 Siemens SOMATOM Confidence (version: syngo CT VB20A)
 Siemens SOMATOM Edge Plus (version: syngo CT VB20A)
 Siemens SOMATOM Force (version: syngo CT VB20A)
 Siemens SOMATOM Perspective
 Siemens SOMATOM go.All (Version VA20A)
 Siemens SOMATOM go.Now (Version VA20A)
 Siemens SOMATOM go.Top (Version VA20A)
 Siemens SOMATOM go.Up (Version VA20A)
 Siemens SOMATOM go.Now (Version: VA30)
 Siemens SOMATOM go.Up (Version: VA30)
 Siemens SOMATOM go.All (Version: VA30)
 Siemens SOMATOM go.Top (Version: VA30)

Siemens SOMATOM go.Sim (Version: VA30)
 Siemens SOMATOM go.Open Pro (Version: VA30)
 Siemens SOMATOM X.cite (Version: VA30)
 Siemens SOMATOM go.Now (Version: VA40)
 Siemens SOMATOM go.Up (Version: VA40)
 Siemens SOMATOM go.All (Version: VA40)
 Siemens SOMATOM go.Top (Version: VA40)
 Siemens SOMATOM go.Sim (Version: VA40)
 Siemens SOMATOM go.Open Pro (Version: VA40)
 Siemens SOMATOM X.cite (Version: VA40)
 Siemens SOMATOM X.ceed (Version: VA40)
 Siemens SOMATOM go.Fit (Version: VA40)
 Siemens Device, SOMATOM Confidence, Version: syngo CT VB30A
 Siemens Device, SOMATOM Definition AS, Version: syngo CT VB30A
 Siemens Device, SOMATOM Definition Edge, Version: syngo CT VB30A
 Siemens Device, SOMATOM Definition Flash, Version: syngo CT VB30A
 Siemens Device, SOMATOM Drive, Version: syngo CT VB30A
 Siemens Device, SOMATOM Edge Plus, Version: syngo CT VB30A
 Siemens Device, SOMATOM Force, Version: syngo CT VB30A
 Siemens Device, SOMATOM Force Velo, Version: syngo CT VB30A
 Siemens Symbia Intevo Excel
 Siemens Symbia Intevo 2
 Siemens Symbia Intevo 6
 Siemens Symbia Intevo 16
 Siemens Symbia T
 Siemens Symbia T2
 Siemens Symbia T6
 Siemens Symbia T16
 Siemens , Device: NAEOTOM Alpha, Version: syngo CT VA50A
 Siemens Healthineers, Product: NAEOTOM Alpha, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM X.ceed, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM X.cite, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.All, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.Fit, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.Now, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.OpenPro, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.Sim, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.Top, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM go.Up, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM Pro.Pulse, Software Version: syngo CT VB10A
 Siemens Healthineers, Product: SOMATOM On.Site, Software Version: syngo CT VB10A
 Toshiba Aquilion (Aquilion Platform)
 Toshiba Aquilion 4 (Aquilion Platform)
 Toshiba Aquilion 8 (Aquilion Platform)
 Toshiba Aquilion 16 (Aquilion Platform)
 Toshiba Aquilion 32 (Aquilion Platform)
 Toshiba Aquilion 64 (Aquilion Platform)
 Toshiba Aquilion CX (Aquilion Platform)
 Toshiba Aquilion ONE (Aquilion Platform)
 Toshiba Aquilion ONE Genesis
 Toshiba Aquilion ONE Vision (Aquilion Platform)
 Toshiba Aquilion PREMIUM (Aquilion Platform)
 Toshiba Aquilion PRIME (Aquilion Platform)
 Toshiba Aquilion RXL (Aquilion Platform)
 Toshiba Aquilion Lightning
 Toshiba Aquilion Super 4
 Toshiba Alexion (Aquilion Platform)
 Toshiba Asteion
 Toshiba Large Bore (Aquilion Platform)
 NeuWave Medical, Inc. Ablation Sys/Discovery RT, v. 3.0.4.451²³

Notes

23 This device is not approved for Query Retrieve

Dental Extra-Oral Radiography

Supported Dental Extra-Oral Radiography Devices

Carestream Dental (Kodak) 9000 Pano (Trophy Platform - DX)
Carestream Dental (Kodak) Ceph-9000 (Trophy Platform - DX)
Carestream Dental (Kodak) Ceph-V3 (Trophy Platform - DX)
Carestream Dental (Kodak) K8000E (Trophy Platform - DX)
Carestream Dental (Kodak) K9000 (Trophy Platform - CT)
Carestream Dental (Kodak) K9500 (Trophy Platform - CT)
Carestream Dental (Kodak) CS-8100 (Trophy Platform TW 6.13.3.2)
Carestream Dental (Kodak) CS-8100 (Trophy Platform TW 6.4)
Carestream Dental (Kodak) CS-9300 (Trophy Platform TW 6.4)
Carestream Dental, CS-9600 (CT, DX), v: 1.0
Cefla North America, Inc MyRay Hyperion MRT²¹
Cefla North America, Inc MyRay Hyperion X5²¹
Cefla North America, Inc MyRay Hyperion X7²¹
Cefla North America, Inc MyRay Hyperion X9²¹
Cefla North America, Inc NewTom Giano²¹
Planmeca Dimax
Planmeca Dimax2
Planmeca ProMax PX-1102
Planmeca ProMax/Romexis 3D CBVT (Code Beam Volumetric Tomography)
Planmeca Romexis ProMax v. 6.2 – (Planmeca Romexis Platform)
Planmeca Proline Pan – (DIMAX3 - Planmeca Romexis Platform)
Planmeca Promax Pan – (DIMAX3 - Planmeca Romexis Platform)
Dentsply Sirona GALILEOS (Code Beam CT)
Dentsply Sirona GALILEOS 3D (Cone Beam CT)
Dentsply Sirona ORTHOPHOS XG 3 Digital Pan (no-Ceph)
Dentsply Sirona ORTHOPHOS XG 3D D3352 Digital Pan (SIDCOM3.0, SIDEXIS XG, GALAXIS)
Dentsply Sirona ORTHOPHOS XG 3D D3352 Digital Pan w/ Ceph (SIDCOM3.0, SIDEXIS XG, GALAXIS)
Dentsply Sirona ORTHOPHOS XG 5 Digital Pan
Dentsply Sirona ORTHOPHOS XG 5 Digital Pan-Ceph
Dentsply Sirona ORTHOPHOS XG Plus Digital Pan 3D ready [2D system] (SIDCOM3.0, SIDEXIS XG)
Dentsply Sirona ORTHOPHOS XG Plus Digital Pan-Ceph 3D ready SIDCOM3.0, SIDEXIS XG)
Dentsply Sirona ORTHOPHOS SL 3D Ai (all s)
Dentsply Sirona ORTHOPHOS SL I (all s)
Dentsply Sirona ORTHOPHOS SL (DCS) (all s)
KaVo Dental Technologies/Instrumentarium Dental Snapshot (Clinview10)
KaVo Dental Technologies/Instrumentarium Dental OP200 (Clinview10)
KaVo Dental Technologies/Instrumentarium Dental OP30 (Clinview10)
KaVo Dental Technologies/Instrumentarium Dental Snapshot (Clinview10)
KaVo Dental Technologies/Soredex CraneX 3D (Clinview10)
KaVo Dental Technologies/Soredex CraneX 3D (Clinview10)
KaVo Dental Technologies/Soredex CraneX Novus (Clinview10)

KaVo Dental Technologies/Soredex CraneX Optime (Clinview10)
KaVo Dental Technologies/Instrumentarium OP300 (Cliniview 10.2)
KaVo Dental Technologies/Instrumentarium OP300 Maxio (Cliniview 10.2)

Notes

21 Requires using NNT/iRYS Workstation w/DICOM module

Dental Intra-Oral Radiography

Supported Dental Intra-Oral Radiography Devices

Carestream Dental (Kodak) RVG 6100 (Trophy Platform - IO)
Carestream Dental (Kodak) RVG 6200 (Trophy Platform - IO)
Cefla North America, Inc MyRay Zen-X Sensor²¹
Cefla North America, Inc MyRay X-Pod Sensor²¹
Cefla North America, Inc MyRay C-U2²¹
Cefla North America, Inc MyRay C-U2 HD²¹
Cefla North America, Inc NewTom VGi²¹
Cefla North America, Inc NewTom VGi Evo²¹
Cefla North America, Inc NewTom 5G²¹
Cefla North America, Inc NewTom GO²¹
Dentsply Sirona IO 2 (Wall-box Ethernet version) Intra Oral Sensors; Sizes 1 & 2
Dentsply Sirona IO 2 Plus Wall Module D3495 (Ethernet version) Intra Oral Sensors; Sizes 1 & 2 (SIDCOM 3.0, SIDEXIS XG)
Dentsply Sirona IO 2 Plus Wall Module D3495 (USB version) Intra Oral Sensors; Sizes 1 & 2 (SIDCOM 3.0, SIDEXIS XG)
Dentsply Sirona XIOS (USB version) Intra Oral Sensors; Size 1 & 2
Dentsply Sirona Schick 33 Sensors; Size 0, 1 & 2, USB Type
Dentsply Sirona Schick Elite Sensors; Size 0, 1 & 2, USB Type
Dentsply Sirona Schick 33 Sensors; Size 0, 1 & 2, USB Type
Dentsply Sirona Schick Elite Sensors; Size 0, 1 & 2, USB Type
Dentsply Sirona XIOS Plus (USB version) Size 1 & 2
Dentsply Sirona XIOS Plus (Ethernet version) Size 1 & 2
Dentsply Sirona USBCam4
Dentsply Gendex DenOptix
Dexis
DEXIS, Device: DEXIS Titanium³⁵, Version: 9.5.1
DEXIS, Device: IXS Size 1³⁵, Version: 9.5.1
DEXIS, Device: IXS Size 2³⁵, Version: 9.5.1
Planmeca DIXI, DIXI2, and DIXI3 Series Detectors
Planmeca Intrascan
Planmeca Multiscan
Planmeca ProSensor (Planmeca Romexis Platform)
Planmeca ProSensor v. 6.2 (Planmeca Romexis Platform)
Planmeca, Product: ProSensor, Version: Romexis 6.3.x
Planmeca, Product: ProMax, Version: Romexis 6.3.x
Planmeca, Product: ProSensor, Version: Romexis 6.4.x
Planmeca, Product: ProMax, Version: Romexis 6.4.x
KaVo Dental Technologies/Soredex Scanora

Notes

21 Requires using NNT/iRYS Workstation w/DICOM module.

Dental VL – Endoscopy
Supported VL – Endoscopy Devices

Carestream Dental (Kodak) K5100 (Trophy Platform - ES)

Medicor Imaging, a division of LEAD Technologies, Inc., Device: MiPACS Dental Enterprise Viewer, Version: 4.3.0.58

Dental Workstations

Supported Dental Workstations

Cefla North America, Inc NNT/iRYS
GE Dental CliniView
Instrumentarium Imaging CliniView
Medicor MIPACS Dental Enterprise PACS
Medicor MIPACS Dental Enterprise PACS, v. 4.2.0.77
Planmeca Dimaxis Pro platform
Dentsply Sirona (Schick Technologies) CDR DICOM
Dentsply Sirona GALAXIS Platform
Dentsply Sirona SIDEXIS XG Platform
Dentsply Sirona SIDICOM 3.0 Platform
Dentsply Sirona SIDEXIS Software for Products (SIDCOM Platform 3.0)
Dentsply Sirona SIDICOM WLS (SIDEXIS XG Platform)
Dentsply Sirona SIDICOM WLS 3.1
Dentsply Sirona SIDEXIS 4
KaVo Dental Technologies/Soredex Clinview 10 Software (Clinview10)
KaVo Dental Technologies/Soredex Scanora 5.0 Software (Clinview10)
Trophy Radiologie RVGui

Digital Subtraction

Supported Digital Subtraction Devices

Philips Integris
Siemens AngioStar
Siemens Polytron

Direct X-Ray

Supported Direct X-Ray Devices

Agfa DX-D 100 (Platform: NX)
Agfa DX-D 300 (Platform: NX)
Agfa DX-D 400 (Platform: NX)
Agfa DX-D 500n (Platform: NX)
Agfa DX-D 600 (Platform: NX)
Agfa DR 400
Agfa DR 600
Agfa Instant DR Retrofit
Canon CXDI
Canon CXDI Control NE
Canon Soltus 500 with CXDI, v. 2.19
Canon, RadPro Mobile 40kw Flex (with CXDI Control NE) (V. 2.17.0.20)
Canon RadPRO Elite XA (w/CDXI Control NE)
Canon RadPRO Elite XM (w/CDXI Control NE)
Canon RadPRO Elite FM1 (w/CDXI Control NE)
Canon RadPRO Mobile 40kW (w/CDXI Control NE)
Canon RadPRO Mobile 40kW Flex (w/CDXI Control NE)
Canon RadPRO Mobile 40kW Flex Plus (w/CDXI Control NE)
Canon RadPRO Omnera 400A (w/CDXI Control NE)
Canon RadPRO Omnera 400T (w/CDXI Control NE)
Canon RadPRO Delinia 200 (w/CDXI Control NE)
Canon RadPRO Soltus 100M (w/CDXI Control NE)

Canon RadPRO Soltus 100 (w/CDXI Control NE)
 Canon RadPRO URS (w/CDXI Control NE)
 Carestream DR7500
 Carestream DRX-1
 Carestream DRX-Evolution
 Carestream DRX-Evolution, v.1.3
 Carestream DRX-Evolution, v.1.5
 Carestream DRX-Evolution Plus, v.1.7
 Carestream DRX-Revolution
 Carestream DRX-Revolution, v:1.5²⁸
 Carestream DRX-Compass, v.1.6²⁸
 Carestream, DRX Revolution Mobile (V. Image View 1.2)
 Carestream DR 3500
 Carestream DR 9500
 Carestream DRX-Mobile Retrofit Kit
 Carestream DRX Excel Plus, v. 1.20
 Carestream DRX-EVO 1.9²³
 Carestream DRX-Mobile 1.9²³
 Carestream DRX-Revo 1.9²³
 Carestream DRX-RevoPlus 1.9²³
 Carestream DRX-EVO PLUS 1.9²³
 Carestream DRX-Ascend v.1.10²³
 Carestream DRX-EVO v.1.10²³
 Carestream DRX-EVO PLUS v.1.10²³
 Carestream DRX-Revo v.1.10²³
 Carestream DRX-RevoPlus v.1.10²³
 Carestream DRX-Compass, v. 1.10²³
 Carestream DRX-InRoom v.1.10²³
 Carestream, DRX-Ascend, v1.11
 Carestream, DRX-Compass, v1.11
 Carestream, DRX-Evolution, v1.11
 Carestream, DRX-Evolution Plus, v1.11
 Carestream, DRX- InRoom, v1.11
 Carestream, DRX-Revo, v1.11
 Carestream, DRX-Revo Plus, v1.11
 Carestream, DRX-Mobile, v1.11
 Carestream, DRX-Revolution Nano, v1.11
 Carestream, DRX-Rise, v1.11
 Carestream, DRX-Transportable, v1.11
 Carestream, DRX-Transportable Lite, v1.11
 Carestream, Device: DRX-Rise, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Compass, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Ascend, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Revo, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Revo Plus, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Nano, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-InRoom, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Evolution, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Evolution Plus, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Transportable Lite, Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Transportable , Software: ImageView, Version: V2.00
 Carestream, Device: DRX-Mobile, Software: ImageView, Version: V2.00
 DEL Medical DelWorks Medical DR System

CMT SmartRAD
Faxitron BioVision
Faxitron PathVision
Faxitron CoreVision
Fuji AcSelerate (Platform: FDX Console)
Fuji Console Advance - DR-ID 300CL
Fuji Console Advance - DR-ID 300CL (Platform: FDX Console)
Fuji D-EVO (Platform: FDX Console)
Fuji Unity
Fuji Velocity
GE ADS_1.0
GE Definium 5000
GE Definium 8000
GE Definium AMX700
GE Discovery XR 650 (Platform Discovery XR)
GE Discovery XR 656 (Platform Discovery XR)
GE Discovery XR656+
GE Discovery XR656HD, v.1.0
GE Discovery XR 650 (Platform Discovery XR)
GE Healthcare Discovery RF180, v. 0.0.3.128
GE Optima 220 (Platform Discovery XR)
GE Optima XR646
GE Optima XR646 HD, v.1.0
GE Optima XR240amx, v. 2.27.123
GE Proteus US
GE Revolution XQ/i
GE Revolution XRd
GE HealthCare, Product: AMX Navigate³⁹, Version: XRAY_MOBILE2.1_4.2.1
Hologic 10004
Hologic 50004
Hologic DROC
Hologic Radex4
Kodak DirectView DR5000
Kodak DirectView DR5100
Kodak DirectView DR7100
Kodak DirectView DR7500
Kodak DirectView DR9000
Konica Minolta CS-2/iQue
Konica Minolta DIRECT DIGITIZER CS-7
Konica Minolta Regius CS-3 (0862)
Konica Minolta/GE AeroDR - CS-7, V: 1.30R02_007
Philips Digital DIAGNOST VR
Philips Digital DIAGNOST
Philips DIGITAL DIAGNOST R3.0 (PLATFORM: EWS0409)
Philips MOBILEDIAGNOST WDR (PLATFORM: EWS0409)
Philips Digital DIAGNOST, v.4.023,28
Philips Digital DIAGNOST, v.4.123,28
Philips Digital DIAGNOST, v.4.223,28
Philips Digital DIAGNOST wDR, v.2.023,28
Philips Digital DIAGNOST wDR, v.2.123,28
Philips Digital DIAGNOST wDR, v.2.223,28
Philips Digital DIAGNOST C90 v.1.123,28
Philips Eleva Workspot 2.0

Philips Essenta Compact
Philips Thoravision
Philips ProxiDiagnost N90 Rel 1.1, v.1.1.1
Philips Medical Systems Netherlands BV, Azurion, R2.2
Philips Medical Systems Netherlands BV, Azurion, R3.0
Philips Medical Systems, Device: Philips Digital Diagnost, Version: Release 4.3
Philips, Product: Radiography 7000 M²⁸, Version: 1.0
Philips, Product: Radiography 7300 C²⁸, Version: 1.0
Samsung GC85A
Samsung GC80
Samsung GU60A
Samsung CF50
Samsung GM60A
Samsung GR40CW
Samsung GM85, v3.03.17
Samsung GM85, v3.06.01 B79
Samsung NeuroLogica, Product: GC85A, Version: v3.08.01
Samsung NeuroLogica, Product: GC85A, Version: v3.09
Samsung NeuroLogica, Product: S-Station Family, Version: v3.09.03 B22
Samsung NeuroLogica, Product: S-Station⁴⁰, Version: v3.11
Siemens AXIOM Multix M
Siemens AXIOM Multix MP
Siemens AXIOM Multix MT
Siemens AXIOM Multix MU
Siemens Cios Alpha, v:VA30
Siemens Cios Spin, v:VA30
Siemens Cios Select, v:VA20
Siemens Mobilett Mira
Siemens Mobilett Mira Max
Siemens Mobilett XP Digital
Siemens Multix Fusion Max
Shimadzu MobileDaRt Evolution⁴
Shimadzu MobileDaRt Evolution EFX
Shimadzu MobileDaRt Evolution SZ
Shimadzu MobileDaRt Evolution EFX SZ
Shimadzu MobileArt Evolution AeroDR
Shimadzu MobileArt Evolution EFX AeroDR
Shimadzu RadSpeed DR⁴
Shimadzu RadSpeed AeroDR
SwissRay ddR Multi-System
SwissRay HD3000 ddRCompact
SwissRay TR4000 (ddR Modulaire xxx/ddR Combi xxx)
SwissRay XG5000 (ddR Combi xxx/ddR Formula xxx)
Toshiba RadRex-i

Notes

- 3 Data is transferred via an intermediary device. Confirm intermediary device was purchased as well.
23 This device is not approved for Query Retrieve

Evidence Creator (Encapsulated PDF)

Supported Encapsulated PDF Devices

Carl Zeiss Meditec-,Product: ATLAS 500²⁸, Version: 1.0

Endosoft (UTECH) Endosoft Systems

Medima Solutions PacsPrinter

NeuroCom International BASIC Balance Master

NeuroCom International BASIC Balance Master (Platform: Balance Manager)

NeuroCom International Balance Master

NeuroCom International Balance Master (Platform: Balance Manager)

NeuroCom International EquiTest

NeuroCom International EquiTest (Platform: Balance Manager)

NeuroCom International EquiTest CRS

NeuroCom International EquiTest CRS (Platform: Balance Manager)

NeuroCom International inVision

NeuroCom International inVision (Platform: Balance Manager)

NeuroCom International PRO Balance Master

NeuroCom International PRO Balance Master (Platform: Balance Manager)

NeuroCom International SMART Balance Master

NeuroCom International SMART Balance Master (Platform: Balance Manager)

NeuroCom International SMART EquiTest

NeuroCom International SMART EquiTest (Platform: Balance Manager)

NeuroCom International SMART EquiTest CRS

NeuroCom International SMART EquiTest CRS (Platform: Balance Manager)

NeuroCom International VSR

NeuroCom International VSR (Platform: Balance Manager)

Film Digitizer

Supported Film Digitizers

Array Corporation AR2905⁴

Hologic, Inc. Digital Now Film Scanner (MG)⁵

Lumisys LS7⁴

Lumisys LS85⁴

Lumisys LS150⁴

Lumisys LS200⁴

Vidar Film Digitizer⁴

Notes

4 Data is transferred via an intermediary device. Confirm intermediary device was purchased as well.

5 Tested as DICOM Object. Test images were not converted to Targa format.

Film Digitizer Software

Supported Film Digitizer Software

Lumisys DI-2000

Mitra Express

Vidar Clinical Express

Frame Grabber

Supported Frame Grabbers

ALI NewPort
Aspect Electronics *See Kodak MIM¹*
Camtronics NT-100¹
Camtronics NT-200¹
DicomIt Information Manager
Foresight Imaging TIMS DICOM SYSTEM
Kodak MIM
Kodak MIMS-100
Kodak MIMS-200
RIS I2000 VICS²

Notes

- 1 No DICOM Modality Worklist capability.
- 2 Poor DICOM Modality Worklist capability.

General Microscopy

Supported General Microscopy Devices

STI Hawaii, Inc. STI-Colposcope

Magnetic Resonance

Supported Magnetic Resonance Devices

Canon Medical Systems, MR Vantage Orian (v.5.0)
GE ADW
GE Discovery MR 450
GE Discovery XR656 HD, v.1.0
GE Discovery MR 750
GE Discovery MR 750W
GE Optima MR430S
GE Optima MR 450
GE Optima MR 450w
GE Signa
GE Signa Artist, v. DV26 ²²
GE Signa HDe
GE Signa HDe 1.5T (14.0) (Platform: Signa HD)
GE Signa HDi
GE Signa HDi 1.5T (14.0) (Platform: Signa HD)
GE Signa HDx
GE Signa HDx 1.5T (14.0) (Platform: Signa HD)
GE Signa HDx 3.0T (14.0) (Platform: Signa HD)
GE Signa Ovation HD (14.0) (Platform: Signa HD)
GE Signa Profile
GE Signa Profile HD (14.0) (Platform: Signa HD)
GE Signa Premier MR, v. RX 28
SIGNA Voyager
SIGNA Explorer
SIGNA Creator
SIGNA Architect
SIGNA Artisan
GE ONI Medical Systems MR 430

GE ONI Medical Systems MSK Extreme
 GE Healthcare, Product: SIGNA Artist, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Premier, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Architect, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Pioneer, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Hero, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Voyager, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Explorer, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA Creator, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA HD, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: SIGNA MR335/MR360/MR380, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Brivo 335, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Optima 360, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Optima MR450w, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Discovery MR750, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Discovery MR750w, Version: SIGNA_LX1.MR30.0_R01_2242.a
 GE Healthcare, Product: Discovery MR450w, Version: SIGNA_LX1.MR30.0_R01_2242.a
 Hitachi AIRIS II
 Hitachi AIRIS Elite
 Hitachi Altaire
 Hitachi Echelon
 Hitachi Echelon Oval
 Hitachi Oasis
 Hitachi Oasis XP
 Marconi Eclipse 1.5T
 Marconi Edge/Asset
 Marconi Edge
 Marconi Vista
 Philips Achieva
 Philips GyroScan
 Philips GyroScanNT
 Philips Gyroscan Intera
 Philips Ingenia
 Philips Intera
 Philips Invivo DynaCad ²⁶
 Philips Invivo DynaCad v4.0 ²⁶
 Philips Invivo DynaCad Breast
 Philips Invivo DynaCad Post Processing Software/Prostate
 Philips Panorama
 Philips, Device: Ingenia Ambition S²⁸, Version: 5.8.1.0
 Philips, Product: Philips MR 5300²⁸, Version: 5.8.1
 Philips, Product: MR 5300²⁸, Version: 11.1.0
 Philips, Product: MR 7700²⁸, Version: 11.1.0
 Philips, Product: Ingenia Ambition X 1.5T²⁸, Version: R11.1
 Philips, Product: Ingenia Ambition S 1.5T²⁸, Version: R11.1
 Philips, Product: Ingenia Elition S 3.0T²⁸, Version: R11.1
 Philips, Product: Ingenia Elition X 3.0T²⁸, Version: R11.1
 Philips, Product: Achieva dStream 3.0T²⁸, Version: R11.1
 Philips, Product: Ingenia Evolution 1.5T²⁸, Version: R11.1
 Philips, Product: Achieva dStream 1.5T²⁸, Version: R11.1
 Philips, Product: DYNACAD²⁸, Version: 5.1
 Philips, Product: MR Systems R57.0, Model: Intera 1.5T²⁸, Version: 5.7
 Philips, Product: MR Systems R57.0, Model: Achieva 1.5T and 3.0T²⁸, Version: 5.7

Philips, Product: MR Systems R57.0, Model: Ingenia 1.5T and 3.0T²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia 1.5T S²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia CX/Achieva dStream 1.5T and 3.0T²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Multiva 1.5T²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Prodiva 1.5T CX²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Prodiva 1.5T CS²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia Ambition S²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia Ambition X²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia Elition S²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Ingenia Elition X²⁸, Version: 5.7
Philips, Product: MR Systems R57.0, Model: Marlin 1.5T²⁸, Version: 5.7

Qlarity Imaging, QuantX, V: 2.0²⁶

Siemens Allegra

Siemens Avanto (*syngoMR*)

Siemens Concerto

Siemens Espree (*syngoMR*)

Siemens GBS III³

Siemens Harmony

Siemens Impact

Siemens MAGNETOM³

Siemens MAGNETOM Aera

Siemens MAGNETOM Amira

Siemens MAGNETOM Skyra

Siemens MAGNETOM Verio

Siemens MAGNETOM –Vision³

Siemens MAGNETOM Skyra^{fit}

Siemens MAGNETOM Avanto^{fit}

Siemens MAGNETOM Prisma^{fit}

Siemens MAGNETOM Prisma

Siemens MAGNETOM Vida

Siemens Sonata

Siemens Symphony

Siemens *syngoMR*

Siemens Trio (*syngoMR*)

Siemens Healthineers , Product: MAGNETOM Sola, Software Version: syngo MR XA51

Siemens Healthineers , Product: MAGNETOM Altea, Software Version: syngo MR XA51

Siemens Healthineers , Product: MAGNETOM Sola Fit, Software Version: syngo MR XA51

Siemens Healthineers , Product: MAGNETOM Viato.Mobile, Software Version: syngo MR XA51

Toshiba Excelart Vantage (MRT-1503)

Toshiba Excelart Vantage Atlas (MRT-1503)

Toshiba Excelart Vantage Titan (MRT-1504)

Toshiba Vantage Elan 1.5T

Toshiba Vantage Titan 1.5T⁴

Toshiba Vantage Titan 3T⁴

Toshiba Vantage Galan 3T

Notes

3 Tested using the DeJarnette Netshare for interfacing. Supported

4 “MPower” platform MR for Toshiba validated

26 Device does not perform MWL, it is a Vista Storage only device.

Mammography

Supported Mammography Devices

CompAI Healthcare, Product: TrueView 100 Pro, Version: R1.0.10
GE Senographe Essential⁵
Faxitron Bioptics Corevision
Faxitron Bioptics BioVision
Faxitron Bioptics PathVision
Hologic Affirm Prone Biopsy System (Version 1.0.2.20)²²
Hologic Brevera/CorLumina, Version 1.0
Hologic Genova (CAD – Secondary Capture)
Hologic Multicare Platinum
Hologic SecureView Diagnostic Mammography Workstation^{5,17}
Hologic Selenia Dimensions (Selenia Dimensions V1.0 platform)⁵
Hologic Selenia Dimensions⁵
Hologic Selenia Dimensions, V1.9¹⁷
Hologic Selenia Dimensions, V1.11.1
Hologic Trident RC Specimen Workstation
Hologic, Product: 3Dimensions, Version: 2.2.1
KUBTEC, Product: Mozart , Version: 11.0.4
Senographe Pristina (Version: NGA_1.20)
Senographe Pristina 3D (Version: NGA_1.20)
Siemens Mammomat Fusion/Inspiration
Siemens Mammomat Inspiration
Siemens Mammomat Revelation¹⁷
Siemens Mammomat Revelation VC10C

Notes

5 Tested as DICOM Object. Test images were not converted to Targa format.

17 DICOM Query/Retrieve – Study/Root Q/R fully supported.

Middleware – Interface Engine

Supported Middleware – Interface Engines

Access AAM
Agfa Impax AS3000¹
Agfa Impax DS3000¹
Agfa Impax NG3000¹
Carl Zeiss Meditec, Inc FORUM
DeJarnette ImageshareCR
DeJarnette MediShare
IMAGE Information Systems iQ-WEBX
IMAGE Information Systems iQ-VIEW
InfiMed GoldOne²
InfiMed PlatinumOne
Foresight Imaging LLC TIMS 2000 SP
Kodak Cluster Server
Medtronics Stealth Station S7
Merge Technologies Examworks
Merge Technologies MVP
MiPACS Storage Server, v. 4.2.0.77
Mitra Agfa PACS Broker
Mitra GE PACS Broker

Mitra PACS Broker
Mitra Picker PACS Broker
PACSGEAR (formerly NAI) DICOM Box
PACSGEAR PacsSCAN
PACSGEAR PacsSCAN Film
VIASYS Vasoguard
Visual-Med DicomPush

Notes

- 1 No DICOM Modality Worklist capability.
- 2 Poor DICOM Modality Worklist capability.

Nuclear Medicine

Supported Nuclear Medicine Devices

ADAC Forte⁴
ADAC Genie¹
ADAC Pegasys (Wks.)²
ADAC Solus⁴
ADAC Vertex⁴
Biodex Medical Systems, Inc Atomlab 960 Thyroid Uptake System
Capintec Captus 3000 (Thyroid Uptake System)
Digirad Corporation Prominence (See Digirad SeeQuanta)
Digirad Corporation SeeQuanta (DICOM Modality Worklist software)⁴
Elscint Expert
GE Discovery NM 530c
GE Discovery NM 570c
GE Discovery NM 630
GE NM/CT 800 Family, v:1.0²³
GE Infinia Camera⁴
GE Hawkeye Camera
GE Starcam
GE Varicam
GE Ventri Camera⁴
GE Xeleris workstation
GE Xeleris 2 Workstation
GE Xeleris 3 Workstation
GE Xeleris 4 Workstation
GE Healthcare, Device: NM/CT 870 DR, Version: 1.003
GE Healthcare, Device: NM/CT 850, Version: 1.003
GE Healthcare, Device: NM/CT 870 CZT, Version: 1.003
GE Healthcare, Device: NM/CT 860, Version: 1.003
GE Healthcare, Device: MyoSPECT, Version: 1.003.542.4
GE Healthcare, Device: MyoSPECT ES, Version: 1.003.542.4
GE Healthcare, Product: NM 830, Version: V1.003
GE Healthcare, Product: StarGuide, Version: V1.003
Marconi IRIX
Marconi Odyssey¹
Marconi P3000⁴
Marconi XP2000⁴
Pinestar Technologies NMIS DMWL⁴
ec² Software Solutions NMIS DMWL⁵
Philips Brightview Gamma Camera⁴
Philips CardioMD⁴
Philips JetStream
Philips Precedence
Siemens ICON workstation¹
Siemens Orbiter
Trionix Research BIAD
Trionix Research BIAD_24
Trionix Research TRIAD

Notes

- 1 No DICOM Modality Worklist capability.
- 2 Poor DICOM Modality Worklist capability.
- 4 Data transferred via an intermediary device. Confirm intermediary device was purchased as well.
5. HL7 interface for MWL only.

23 This device is not approved for Query Retrieve

Nuclear Medicine - SPECT

Supported Nuclear Medicine – SPECT

GE Brivo NM 615 (NM 600 Platform)
GE Discovery NM 630 (NM 600 Platform)
GE NM 600 Series (Platform)
GE Discovery NM/CT 670
GE Discovery NM/CT 670 Pro
GE Discovery NM/CT 670 ES
GE Optima NM/CT 640 (NM 600 Platform)
GE Vision POWERstation
Philips Brightview XCT
Philips Extended Brilliance Workplace
Siemens e.soft
Siemens Symbia E (eSoft Platform)
Siemens Symbia eSoft Platform (v8.5)
Siemens Symbia S (eSoft Platform)
Siemens Symbia T (eSoft Platform)
Siemens Symbia T2 (eSoft Platform)
Siemens Symbia T6 (eSoft Platform)
Siemens Symbia T16 (eSoft Platform)
Siemens Symbia E (syngo MI Applications platform)
Siemens Symbia Evo (syngo MI Applications platform)
Siemens Symbia Evo Excel (syngo MI Applications platform)
Siemens Symbia S (syngo MI Applications platform)
Siemens, Product: Symbia Pro.specta X3, Software: SPECTsyngo, Version: VA10
Siemens, Product: Symbia Pro.Specta Q3, Software: SPECTsyngo, Version: VA10
Siemens, Product: Symbia Pro.Specta x7, Software Version: SPECTsyngo VA10B
Spectrum-Dynamics DSPECT

Ophthalmic Photography

Supported Ophthalmic Photography Devices

Canon imageSPECTRUM Review v 5.3
Canon CF1 (MYD) w/imageSPECTRUM Capture Plus v.5.0
Canon CF1 (MYD) w/imageSPECTRUM Review v.5.1
Canon CR1 Camera (non-MYD)
Canon CR1 Mark II (non-MYD)
Canon CR2 (Platform-Retinal Imaging Control 3.2)
Canon CR2 Plus (Platform-Retinal Imaging Control 3.2)
Canon CX1 (Platform-Retinal Imaging Control)
Canon EyeQ Pro
Canon CR2 (Platform-Retinal Imaging Control 4.1)
Canon CR-2 Plus AF (Platform-Retinal Imaging Control 4.1)
Canon CX-1 (Hybrid) w/imageSPECTRUM Capture Plus Version 5.0
Canon CX-1 (Hybrid) w/imageSPECTRUM Review Version 5.1
Canon CR2 w/imageSPECTRUM Capture Plus Version 5.0
Canon CR-2 Plus AF w/imageSPECTRUM Capture Plus Version 5.0
Canon CR-2 Plus AF w/imageSPECTRUM Review Version 5.1
Canon CR-2 AF w/imageSPECTRUM Review Version 5.1
CR-2 Plus AF with imageSPECTRUM Capture Plus, v.6.0

Canon Medical Systems, imageSPECTRUM Capture Plus Version 6 with the CR-2 AF, v6.0
 Canon Medical Systems, imageSPECTRUM Capture Plus with the CX-1, v6.0
 Canon Medical Systems, imageSPECTRUM - Capture Plus with the CR-2 Plus AF, v6.0
 Canon Medical Systems, Device: CR-2 AF, Version: 4.4.1.9
 Canon Medical Systems, Device: CR-2 Plus AF, Version: 4.4.1.9
 Canon Medical Systems, Device: CX-1, Version: 4.4.1.9
 Carl Zeiss Meditec, Inc. Cirrus photo, 2.0.2
 Carl Zeiss Meditec, Inc. CLARUS 500, Version 1.0.x (1.0.4/1.0.5)^{18,24}
 Carl Zeiss Meditec, Inc. CLARUS 500/700 Instrument and Review, v:1.1.1
 Carl Zeiss Meditec, Inc. CLARUS 700, v.1.2
 Carl Zeiss Meditec, Inc. CLARUS 500/700 Version 1.1.4.x
 Carl Zeiss Meditec, Inc. IOL Master 500^{15,18}
 Carl Zeiss Meditec, Inc. IOL Master 700
 Carl Zeiss Meditec, Inc. IOL Master 700, v. 1.88²⁹
 Carl Zeiss Meditec, Inc. IOL Master 700, v. 1.90²⁹
 Carl Zeiss Meditec, Inc. VISUCAM
 Carl Zeiss Meditec, Inc. VISUPAC
 Carl Zeiss Meditec, Inc. FORUM, v. 4.2.x
 Carl Zeiss Meditec, Inc. FORUM²⁷
 Carl Zeiss Meditec, Inc. Photo 600/800 w/FORUM
 Carl Zeiss Meditec, Inc. HFA3/FORUM/Glaucoma Version 1.6/4.2.x/3.5.2.x
 Carl Zeiss Meditec, Inc. Atlas 9000, V: 3.0.2
 Carl Zeiss Meditec, Device: SL Imaging, Version: 2.1.3
 Carl Zeiss Meditec, Product: Cirrus 500 Instrument and Review Station, Software Version: CIRRUS HD-OCT 11.7
 Carl Zeiss Meditec, Product: Cirrus 5000 Instrument and Review Station, Software Version: CIRRUS HD-OCT 11.7
 Carl Zeiss Meditec, Product: Cirrus 6000 Instrument and Review Station, Software Version: CIRRUS HD-OCT 11.7
 CenterVue EIDON (Confocal Scanner)
 CenterVue SpA DRSplus, v. 1.5.3
 Chace and Associates Technologies, LLC ZPIP
 Escalon Ophthavision
 Estenda JVN
 Haag Streit EyeCap
 Haag Streit Octopus 600 w/Eyesuite
 Haag Streit Octopus 900 w/Eyesuite
 Haag Streit Lenstar LS 900 w/Eyesuite
 Haag Streit BX900 w/Eyesuite
 Haag Streit BI900 w/Eyesuite
 Haag Streit Imaging Module 900 w/Eyesuite
 Haag Streit Imaging Module 600 w/Eyesuite
 Haag Streit Fundus module 300 w/Eyesuite
 ifa united I-tech Inc ifa ecPACS
 Heidelberg Engineering HEYEX 2, v. HEYEX 2
 Kowa VKDICOM
 Kowa VK-2
 Medflow oiCapture¹⁵
 Medflow MFCapture
 Nidek 3Dx
 Nidek (PACSPLUS) AFC 210/230 (PPdicom300E)
 Nidek OICAPTURE
 OCULUS Optikgeraete GmbH. Pentacam
 OIS (Merge Healthcare) EyeStation Q5 (Platform: OIS Winstation)
 OIS (Merge Healthcare) EyeStation Q5R (Platform: OIS Winstation)
 OIS (Merge Healthcare) EyeStation Z (Platform: OIS Winstation)
 OIS Winstation (Platform)
 Optos Optos Retinal Scanner

Optos P200T (Suzuka), Version: S90607-1.16
Optos P200DTx (Suzuka), Version: S90607-1.16
Optos P200TE (Suzuka), Version: S90607-1.16
Optos P200TxE (Suzuka), Version: S90607-1.16
Optos California (P200DTX), Nurburg 2.4.3 (Device software); Vantage 3.4
Optos P200T Daytona, Nurburg: 1.6 (Device software); Vantage 3.4
Optos P200T Daytona Plus, Vantage: 3.4.
Optos UWFPRIMARY, v. 3.4
Optos, Product: Silverstone (model number P200TXE), Software: Suzuka , Version: 1.13.0.82 (part number S90607)
Optos, Product: Suzuka P200T
Orion Imaging EyeCap
RadInfo Systems oiCapture¹⁵
Topcon IMAGEnet
Topcon Eyeroute
Topcon EZ Lite2
Topcon IMAGEnet 5
Topcon, Device: 3D-OCT1 Maestro 2, Software: IMAGEnet 6, Version: 2.5
Topcon, Device: TRC NW 400, Software: IMAGEnet 6, Version: 2.5
Topcon, Product:TRC NW 500, Software: IMAGEnet 6, Version: 2.52
Topcon, Product: EzLite 2, Version: v1.4
URAL Telekom Ltd. NewVision Fundus, v.4
Visual Pathways Aris 70
Visual Pathways Aris 70 VigilAnt
Visual Pathways Aris 110
Visual Pathways RSVP
Volk Optical Inc. PictorWorkstation (v. 4.05)

Notes

15 Does not support OP SOP Classes.
18 DICOM Encapsulated PDF only

Ophthalmic Ultrasound

Supported Ophthalmic Ultrasound Devices

Ellex Medical Pty Ltd Eyecubed
Ophthalmic Technologies, Inc. OTI-SCAN
Ophthalmic Technologies, Inc. OTI-SCAN 1000
Ophthalmic Technologies, Inc. OTI-SCAN 2000

Ophthalmic Workstation

Supported Ophthalmic Data Management Workstation

Haag-Streit AG EYESUITE
Topcon Synergy
Topcon 3D-OCT¹⁸
Carl Zeiss Meditec, Inc. Glaucoma Workplace, v 3.1
Carl Zeiss Meditec, Inc. Glaucoma Workplace v.3.5.2.x

Optical Coherence Tomography (Ophthalmology use only)

Supported Ophthalmic Coherence Tomography Devices

Carl Zeiss Meditec, Inc. Cirrus HD-OCT 400/4000¹⁸
Carl Zeiss Meditec, Inc. Cirrus HD-OCT 500/5000¹⁸
Carl Zeiss Meditec, Inc. Cirrus HD-OCT Ver 9.0¹⁸

Carl Zeiss Meditec, Inc. Cirrus HD-OCT Ver 10.0¹⁸
Carl Zeiss Meditec, Inc. Cirrus 500/5000 Instrument and Review, v.9.5.1
Carl Zeiss Meditec, Inc. Cirrus 500/5000 Instrument and Review Station with SW Version 11.5¹⁸
Carl Zeiss Meditec, Inc. Cirrus 500/5000/6000 Instrument and Review Station with SW Version 11.5.1¹⁸
Carl Zeiss Meditec, Inc. Glaucoma Workplace, v 3.1
Carl Zeiss Meditec, Inc., Device: CIRRUS HD-OCT 500, Version: 11.6.x.x
Carl Zeiss Meditec, Inc., Device: CIRRUS HD-OCT 500, Version: 11.6.x.x
Carl Zeiss Meditec, Inc., Device: CIRRUS HD-OCT 500, Version: 11.6.x.x
Heidelberg Engineering Spectralis HRA + OCT
Heidelberg Engineering Spectralis OCT
Ophthalmic Technologies, Inc. OCT Ophthalmoscope
Optovue XR Avanti w/DICOM Module 2.1+
Optovue iCAM w/DICOM Module 2.1+
Optovue iVue w/DICOM Module 2.1+
Optovue RTVue w/DICOM Module 2.1+
Optovue iScan w/DICOM Module 2.1+

Notes:

18 DICOM Encapsulated PDF

Other

Other Supported Devices

DesAcc, Inc. Health Data Explorer
MeVis Medical Solutions AG (Veolity) v.1.4¹
Smith-Nephew 660IMS (works with Integration Broker)
Smith-Nephew Integration Broker
VASAMED - PAD-IQ, 3.05

Notes:

Image Reading Software¹

PET

Supported PET Devices

CTI PET
GE Healthcare, Product: OMNI Legend, Version:1.0
Philips Extended Brilliance Workspace
Philips Gemini GXL
Philips Gemini TF
Philips Astonish TF 16
Siemens Biograph6
Siemens Biograph40
Siemens Biograph64
Siemens mi.CT (Biograph 64)

PET/CT

Supported PET/CT Devices

GE Discovery 610 (Platform: Discovery 2006)
GE Discovery 710 (Platform: Discovery 2006)
GE Discovery PET/CT
GE Discovery MI
GE Discovery MI DR
GE Optima 560 (Platform: Discovery 2006)
GE Healthcare, Product: Discovery IQ, Software version: pet_odyssey5.21
Philips Healthcare Ingenuity PET/CT

Philips Healthcare Vereos PET/CT
Philips Medical Systems, Device: Philips Ingenuity, Version: 4.1.10
Philips Medical Systems, Device: Philips Brilliance64, Version: 4.1.10
Siemens Biograph mCT
Siemens Biograph mCT Flow
Siemens Biograph TruePoint
Siemens Biograph Horizon
Siemens Biograph16_Horizon 4R, VJ30B³²
Siemens Biograph Vision (PT/CT), v.VG75A
Siemens Biograph Vision (PT/CT), v.VG76A
Siemens Biograph Vision (PT/CT), v.VG80A³²
Siemens *syngo* MI.PET/CT 2012A
Siemens m.CT (Platform syngo ml.PET/CT 2011A)
Siemens Somatom Definition AS (Platform syngo ml.PET/CT 2011A)
Siemens Somatom Definition Flash (Platform syngo ml.PET/CT 2011A)
Siemens *syngo* MI.PET/CT 2012A
Toshiba Celesteion

PET/MR

Supported PET/MR Devices

Siemens Biograph mMR

Q/R – 3D Reconstruction Workstation

Supported Query/Retrieve 3D Reconstruction Workstations

TeraRecon Aquarius¹⁷
Vital Imaging Vitrea 3D¹⁷
Voxar 3D¹⁷

Notes

¹⁷ DICOM Query/Retrieve – Study/Root Q/R fully supported.

Q/R – Planning Workstation

Supported Query/Retrieve Planning Workstations

Bronchus Technologies Lungpoint Virtual Bronchoscopic Navigation System¹⁶
OrthoView Workstation¹⁷
Voyant Health (formally Orthocrat) TraumaCad (Orthopedic Templating)¹⁷

Notes

¹⁶ DICOM Query/Retrieve – Performs C-Find SCU and C-Move SCU roles only.
¹⁷ DICOM Query/Retrieve – Study/Root Q/R fully supported

Q/R – Routers

Supported Query/Retrieve Routers

Intelrad IntelePACS
Medweb Server (MAG*3.0*66 only)¹⁶
MIM Software, Product: MIM¹⁶, Version: 7.3
ScImage Picom Enterprise (MAG*3.0*116 v3.3)¹⁶
ScImage Picom Enterprise (MAG*3.0*66 only)¹⁶

Notes

16 DICOM Query/Retrieve – Performs C-Find SCU and C-Move SCU roles only.

Q/R – Workstations

Supported Query/Retrieve Workstations

Agfa Impax 6.5 (cPACS)¹⁶
Agfa Enterprise Imaging, v. 8.1
Bayer Radimetrics Enterprise Platform¹⁷
GE AWS¹⁷
GE Healthcare, Xeleris, Version: 5.0
GE Healthcare, Product: Xeleris, Version: 4.1
Marconi VoxelQ¹⁷
Medimage DICOMLink¹⁶
Medicom Technologies Inc, Medicom Imagex, v2.1.68
MIM Software, v6.8¹
MVISUM, Inc. Cardiology Communications System – DICOM Interface¹⁶
NordicNeuroLab nordicBrainEx¹
Perkins EZ PIC HD
Philips Medical System 3D Roadmap R2 (Platform: XtraVision R8)¹⁷
Philips Medical System Allura 3D-CA R3 (Platform: XtraVision R8)¹⁷
Philips Medical System Allura 3D-RA R6 (Platform: XtraVision R8)¹⁷
Philips Medical System Allura XperCT R2 (Platform: XtraVision R8)¹⁷
Philips Medical System Allura XperGuide R2 (Platform: XtraVision R8)¹⁷
Philips Medical System CTO Navigator (Platform: XtraVision R8)¹⁷
Philips Medical System CTTtrueView R2 (Platform: XtraVision R8)¹⁷
Philips Medical System MultiModality Roadmap R1 (Platform: XtraVision R8)¹⁷
Philips Medical System StentBoost R2 (Platform: XtraVision R8)¹⁷
Philips Medical System StentBoost R3 (Platform: XtraVision R8)¹⁷
Philips Medical System SmartCT v.1.0 (Interventional Workspot)
Philips Medical System SmartPerfusion v. 1.1 (Interventional Workspot)
Philips Medical System VesselNavigator v.1.2 (Interventional Workspot)
Philips Medical System StentBoost v. 4.4 (Interventional Workspot)
Philips Medical System HeartNavigator v.3.2 (Interventional Workspot)
Philips Medical System EP navigator v. 5.4 (Interventional Workspot)
Philips Medical System CT TrueView v.2.0.40 (Interventional Workspot)
Philips Medical System AneurysmFlow v.1.2 (Interventional Workspot)
Philips Medical System 3D-RA v.6.6 (Interventional Workspot)
Philips Medical System 3D-CA v.3.0.40 (Interventional Workspot)
Philips Medical Systems, Application: StentBoost R4.4, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: 3D-CA R3.0, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: 3D-RA R6.6, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: XperGuide R1.3, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: XperCT Dual R3.5, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: EmboGuide R1.2, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Application: 3D Roadmap R1.3, Manufacturer's Model Name: Interventional Workspot, Version: 1.6
Philips Medical Systems, Interventional Workspot, Version: 1.6
Philips Medical Systems, MR-CT Roadmap R1.3, Version:1.6
Philips Healthcare Intellispace Portal 6.0
Philips Healthcare Intellispace Portal 7.0
Philips Healthcare Intellispace Portal 8.0
Philips Healthcare Intellispace Portal 8.0
Philips Healthcare Intellispace Portal, V: ISP11 RFD2¹
Philips Healthcare Intellispace Portal Workstation, v. ISP11¹
Phillips, Product: Intellispace Portal Workstation, Software Version: ISP12
Philips Healthcare DynaCAD Lung¹
Philips, Product: Advanced Visualization Workspace V15, Version: v15
Philips, Product: Advanced Visualization Workspace V15 for CT Applications, Version: v15

Philips, Product: Advanced Visualization Workspace V15 for MR Applications, Version: v15
 Philips, Product: Advanced Visualization Workspace V15 for Multi-Modality Applications, Version: v15
 Philips, Product: Advanced Visualization Workspace V15 for NM Applications, Version: v15
 Philips, Application: 3D Roadmap R1.5, Manufacturer's Model Name: Interventional Workspot, Version: 1.8
 Philips, Application: 3D-RA R6.8, Manufacturer's Model Name: Interventional Workspot, Version: 1.8
 Philips, Application: AneurysmFlow R1.4, Version: 1.4
 Philips, Application: EP navigator R5.6, Version: 5.6
 Philips, Application: EmboGuide R1.4, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: CT TrueView R2.0, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: SmartCT R1.3, Platform: Interventional Workspot, Version: 1.3
 Philips, Application: StentBoost 4.6, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: MR-CT Roadmap R1.5, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: XperGuide R1.5, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: XperCT Dual R3.7³⁸, Platform: Interventional Workspot, Version: 1.8
 Philips, Application: VesselNavigator R1.4³⁸, Platform: Interventional Workspot, Version: 1.4.5
 Philips, Application: HeartNavigator R3.4³⁸, Platform: Interventional Workspot, Version: 3.4
 Philips, Application: 3D-CA R3.0³⁸, Platform: Interventional Workspot, Version: 1.8
 Philips, Product: Interventional workstation³⁸, Version: 1.8
 Philips, Product: Vue PACS³⁸, Version: 12.2.8
 Segami Oasis
 Siemens *syngo* MMWP (Multi-Modality Workplace)¹⁷
 Siemens *syngo.via*¹⁷
 Siemens *syngo* X-Leonardo
 Siemens *syngo* X Workplace
 Siemens *syngo* X Workplace, VD20
 Siemens *syngo* X Workplace, VD30
 Vital Imaging Vitrea Enterprise Suite V1.2¹⁷

Notes

- 1 No DICOM Modality Worklist capability.
- 16 DICOM Query/Retrieve – Performs C-Find SCU and C-Move SCU roles only.
- 17 DICOM Query/Retrieve – Study/Root Q/R fully supported. Formerly named EXPOSURE

Radio Fluoroscopy and X-Ray Angiography
Supported Radio Fluoroscopy and X-Ray Angiography Devices

Canon Medical Alphenix, v: V8.3 SP0001
 Canon Medical Ultimix-I, v: 1.60.04
 Canon Medical Systems,Product: HDR-08A, Model: Ultimix-i, Version: v2.00
 Canon Medical Systems,Product: HDR-08A, Model: Kalare, Version: v2.00
 CMT SmartSpot PrimaX
 Dornier MedTech Compact Delta II (UIMS platform)
 Dornier MedTech Doli-S II (UIMS platform)
 Dornier MedTech Opus II (UIMS platform)
 Dornier MedTech Genesis (UIMS platform)
 Dornier MedTech Gemini (UIMS platform)
 Dornier MedTech UIMS Platform 4.x
 GE DLX
 GE DRS
 GE LCA-DLX Advantx Angio
 GE Mini View 6800
 GE OEC 2800 UroVIEW
 GE OEC 9800
 GE OEC 9900

GE OEC Elite
 GE Precision 500D (Platform Precision 500)
 GE Precision 500D w/ Flashpad (Platform Precision 500)
 GE Precision 600FP, V:1.53.01
 GE Precision RXi¹⁷
 GE Precision MPi
 GE Innova IGS 520
 GE Innova IGS 530
 GE Innova IGS 540
 GE Innova IGS 620
 GE Innova IGS 630
 GE Discovery IGS 730
 GE Discovery IGS 740
 GE Healthcare Discovery RF180, v. 0.0.3.1²⁸
 GE Healthcare, GE OEC 3D, v3.0.0.2351
 GE Healthcare, Product: OEC 3D²⁸, Version: 3.0.0.4532
 GE Healthcare, Product: GE OEC Elite²⁸, Version: 3.1.0.0689
 GE Healthcare, Product: Allia IGS3, Version: DL_22.x.x
 GE Healthcare, Product: Allia IGS5, Version: DL_22.x.x
 GE Healthcare, Product: Allia IGS7, Version: DL_22.x.x
 GE Healthcare, Product: Allia IGS70R, Version: DL_22.x.x
 GE HealthCare, Product: OEC Elite²⁸, Version: 3.1.0.1426
 GE HealthCare, Product: OEC 3D²⁸, Version: 3.1.0.6894
 Hologic InSight
 Hologic InSight Fluoroscanner C-Arm
 InfiMed, Inc. i5 DR (Platform: i5)⁹
 InfiMed, Inc. i5 DSA (Platform: i5)⁹
 InfiMed, Inc. i5 RF (Platform: i5)⁹
 InfiMed, Inc. iCynergy Digital System (Platform: i5)⁹
 InfiMed, Inc. Nexus (Platform: i5)⁹
 Liebel-Florsheim Hydrovision (DR-Plus)⁴
 Medtronic Navigation O-Arm
 OrthoScan HD
 OrthoScan HD with Flat Detector
 OrthoScan HD with Flat Detector - OverRotation
 OrthoScan FD Pulse
 OrthoScan Mobile DI (MDI)
 OrthoScan Mini C-Arm, v. 1.0.2.40
 OXOS Medical, Micro C M01, micro-c-1.0
 PAUSCH Medical GmbH Uroview FD (V. 8.0.15.15)
 Philips Allura XPER FD20
 Philips Azurion v.2.1
 Philips BV Pulsera²
 Philips BV Pulsera (Platform BV Family)
 Philips DataMediatOR CombiDiagnost R90 Rel 1.1.2
 Philips DSI 4.X²
 Philips DSI 5.X²
 Philips EasyDiagnost (DSI 5.X)²
 Philips EasyDiagnost Eleva^{4,9}
 Philips EDI (Extended Digital Imaging)^{4,9}
 Philips Integris
 Philips Integris Allura⁹
 Philips Integris Allura XPER FD10 Cardio⁹

Philips Juno DRF
 Philips MultiDiagnost Eleva^{4,9}
 Philips OmniDiagnost Eleva^{4,9}
 Philips UroDiagnost Eleva^{4,9}
 Philips Veradius (Platform BV Family)
 Philips ViewForum Workstation
 RadPro 40kw Mobile X-Ray System (Canon CXDI)⁴
 Shimadzu DAR8000
 Shimadzu FluoroSpeed (DAR8000)
 Shimadzu Sosialvision Versa (DAR8000)
 Shimadzu SosialVision G4 [with Shimadzu DR-300]
 Shimadzu SosialVision G4 [with Shimadzu DR-300 and Canon CXDI Control NE]
 Shimadzu SosialVision G4 [with Shimadzu DR-300 and Konica Minolta DIRECT DIGITIZER CS-7)
 Shimadzu Digitex Pro
 Shimadzu Corporation, Product: DR-300, Model: SONIALVISION G4³⁹, Software Version: REF Station v1.22.06
 Shimadzu Corporation, Product: DR-300, Model: FluoroSpeed X1³⁹, Software Version: REF Station v1.22.06
 Siemens Arcadis
 Siemens Arcadis C-Arm
 Siemens Artis (Remote Workstation)
 Siemens Artis one
 Siemens Artis zee
 Siemens Artis zee, VD12
 Siemens Artis zeego
 Siemens Artis zeego, VD12
 Siemens Artis Ceiling Zee YMAT
 Siemens Artis Q
 Siemens Artis Q, VD12
 Siemens Artis Q.zen
 Siemens Artis Q.zen, VD12
 Siemens Artis pheno
 Siemens ARTIS icono, VE20
 Siemens ARTIS icono / pheno, VE20
 Siemens Axiom Aristos (FD-X platform)
 Siemens Axiom Aristos MX
 Siemens Axiom Aristos VX/TX
 Siemens Axiom Artis (FD-X platform)
 Siemens Axiom Luminos
 Siemens Cios Alpha
 Siemens Cios Fusion
 Siemens Cios Connect
 Siemens Fluorospot (Platform: Fluorospot Compact VC10)
 Siemens Fluorospot Compact (Platform: Fluorospot Compact VC10)
 Siemens Fluorospot (Platform: Fluorospot Compact VG12)
 Siemens Luminos Agile
 Siemens Luminos Agile Max
 Siemens Luminos Agile Max, VF10
 Siemens Luminos dRF
 Siemens Luminos dRF Max
 Siemens Luminos dRF Max, VF10
 Siemens Luminos Fusion
 Siemens Multitom Rax
 Siemens Siremobil Compact L (uses Memoskop DICOM Bridge)
 Siemens UROSKOP Omnia

Siemens UROSKOP Omnia Max
 Siemens Uroskop Omnia Max, VF10
 Siemens Ysio (Platform: Fluorospot Compact VC10)
 Siemens Ysio, Multi Detector (Platform: Fluorospot Compact VC10)²⁰
 Siemens Ysio Max
 Siemens Ysio Max, VF10
 Toshiba DFP-2000A/AS
 Toshiba DFP-8000 (Infinix)
 Toshiba EPS-Plus (w/ADR-1000A)
 Toshiba Kalare (w/ADR-1000A)
 Toshiba Ultimax (w/ADR-1000A)
 Canon Medical Ultimax-i (w/ HDR-08A)
 Villa Sistemi Medicali Villa Apollo DRF
 Ziehm Imaging 7000Plus⁴
 Ziehm Imaging 8000⁴
 Ziehm Imaging NetPort (Platform Netport)
 Ziehm Imaging NetPort (workstation)
 Ziehm Imaging Vario 3D⁴
 Ziehm Imaging Ziehm Solo (Platform Netport)
 Ziehm Imaging Ziehm Vision (Platform Netport)
 Ziehm Imaging Ziehm Vision FD (Platform Netport)
 Ziehm Imaging Ziehm Vision FD Vario 3D (Platform Netport)
 Ziehm Imaging Ziehm Vision R (Platform Netport)
 Ziehm Imaging Ziehm Vision RDF (Platform Netport)
 Ziehm Imaging Ziehm Vision Vario 3D(Platform Netport)
 Ziehm Vision: RFD Family (Version: 7.05)
 Ziehm Imaging Vista⁴
 Ziehm Imaging VistaPlus⁴
 Ziehm Orthoscan, Product: TAU³⁹, Software Version: 2.0.7

Notes

2 Poor DICOM Modality Worklist capability.
 4 Data transferred via an intermediary device. Confirm intermediary device was purchased as well.
 9 Automatic measurements may not be possible; manual calibration may be required.
 17 DICOM Query/Retrieve – Study/Root Q/R fully supported.

Secondary Capture

Supported Secondary Capture Devices

Abbott (Manufacturer: Coroventis Research AB), Product: CoroFlow Cardiovascular System, Version: 3.6
 Aranz Medical Limited SilhouetteLink
 Bar-Code DICOMVIDEO
 Biodex Medical Systems, Inc Atomlab 960 Thyroid Uptake System
 Echosens, Product: FibroScan, Version: 4.1.5
 E-Web Information Co. DICOM Engine
 GE ViewPoint¹
 GE HealthCare, Product: enCORE, Version: 18.50
 Global Media Group, LLC CapSure CONI
 Global Media Group, LLC CapSure Vista
 Global Media Group, LLC CapSure Version 2.1.6.1
 Global Media Group, LLC CapSure Version 2.1.6.3
 Global Media Group, LLC CapSure Version 2.1.6.4
 Global Media Group, LLC CapSure Version 2.1.6.5³¹

Global Media Group, LLC CapSure Version 3.0
GlobalMedia Group, LLC CapSure Version 3.2
Heidelberg Engineering(Manufacturer : MedicalCommunications GmbH) , Product: HEYEX 2, Version: 2.6.4
Hyperion Medical Technologies, Product: Smart ABI Plus, Version: v4.0.0
Koven Technology, Product: Koven PAUS 2.0, Version: v1.0.1
Laborie Urodynamics System²
Life-Tech, Inc. Urodynamics System
Life-Tech, Inc. Urodynamics System V/IV
Perimed, Product: PSW ExM, Version: v5.6.0
Philips, Product: Dynamic Coronary Roadmap^{1,41}, Platform: IGT, Version: 3.1
Philips, Product: Dynamic StentBoost Live^{1,41}, Platform: IGT, Version: 2.1
Rapid AI, Product: Rapid, Version: 5.7
Rapid AI, Product: Rapid ICH, Version: v3.0
Rapid AI, Product: Rapid Hyperdensity, Version: v1.0
Rapid AI, Product: Rapid PE, Version: v1.0
Rapid AI, Product: Rapid NCCT Stroke, Version: v1.0
Rapid AI, Product: Rapid, Version: v5.9
Riverain Medical Softview (Bone Suppression SW)
Sapheneia Clarity Server (Router)
Siemens healthneers, Product: syngo Dynamics, Version: VA41
TeraRecon, Product: AQNET, Software: Intuition Viewer, Version: 4.7
Topcon(Manufacturer : Visia Imaging S.r.l.), Product: CA-800, Version: 1.6

Notes

1. No DICOM Modality Worklist capability
2. Uses PACScan DICOM Interface

Secondary Capture (Injection Dosage Capture)

Supported Secondary Capture Devices

Bayer or Medrad Connect_PACS
Bayer/Medrad Certegra v.4.70

Surgical Workstation

Supported Surgical Workstations

ImageStream nCare
Infinite Software Solutions, Inc MD Reports
MedXchange CMS9000

Ultrasound

Supported Ultrasound Devices

Accutome 4 Sight
Acuson 128-XP⁵
Acuson KinetDX⁵
Acuson Sequoia^{5,9}
Siemens (Acuson) Sequoia^{5,9}
Acist Medical HDi IVUS
ALI NewPORT

ATL 3000^{1,5}
 ATL 5000⁵
 Bard Access Systems Site-Rite 8 Ultrasound System
 BK Ultrasound (formerly BK Medical) Flex Focus 1202⁹
 BK Ultrasound (formerly BK Medical) ProFocus 2202
 BK Ultrasound BK2300/3000/5000
 Carestream TouchPrime
 Canon Medical Systems, Model: TUS-AI700²⁸, Product: Aplio iSeries, Version: 6.5
 Canon Medical Systems, Model: TUS-AI800²⁸, Product: Aplio iSeries, Version: 6.5
 Canon Medical Systems, Model: TUS-AI900²⁸, Product: Aplio iSeries, Version: 6.5
 Canon Medical Systems, Model: TUS-AI700²⁸, Product: Aplio iSeries, Version: 7.0
 Canon Medical Systems, Model: TUS-AI800²⁸, Product: Aplio iSeries, Version: 7.0
 Canon Medical Systems, Model: TUS-AI900²⁸, Product: Aplio iSeries, Version: 7.0
 CorVascular Falcon
 Digisonics DDI (Digisonics DICOM Implementation)¹⁷
 Esaote (Biosound) MyLab20 Plus^{5,9}
 Esaote (Biosound) MyLab25^{5,9}
 Esaote (Biosound) MyLab30^{5,9}
 Esaote (Biosound) MyLab40^{5,9}
 Esaote (Biosound) MyLab50^{5,9}
 Esaote (Biosound) MyLab60^{5,9}
 Esaote (Biosound) MyLab70^{5,9}
 Esaote (Biosound) MyLab70 XVG^{5,9}
 Esaote (Biosound) MyLab Class C (Platform: Mylab v4.1; v4.3)^{5,9}
 Esaote (Biosound) MyLab Five^{5,9}
 Esaote (Biosound) MyLab One (Platform: Mylab v4.1; v4.3)^{5,9}
 Esaote (Biosound) MyLab Twice (Platform: Mylab v4.1; v4.3)^{5,9}
 Exact Imaging, ExactVu v2.9, v2.9.0.47
 FUJIFILM Sonosite, Product: SONOSITE LX, Version: 100.80.210.064
 FUJIFILM, Product: SONOSITE ST, Version: 100.80.301.083
 GE Healthcare BT18 Voluson, v18
 GE Healthcare (U-SYSTEMS) Invenia ABUS
 GE Healthcare Echopac⁹
 GE Healthcare LOGIQ7⁵
 GE Healthcare LOGIQS7
 GE Healthcare LOGIQ 400⁵
 GE Healthcare LOGIQ 500⁵
 GE Healthcare LOGIQ 700⁵
 GE Healthcare LOGIQ Book⁹
 GE Healthcare LOGIQ E⁹
 GE Healthcare LOGIQ E9⁹
 GE Healthcare LOGIQ E R9.0.3F
 GE Healthcare LOGIQ E10 R1
 GE Healthcare LOGIQ E10, R2.5.2
 GE Healthcare LOGIQ E10, Version: R3.1.2
 GE Healthcare LOGIQ E10s, Version: R3.1.2
 GE Healthcare LOGIQ E11, Version: R3.1.2
 GE Healthcare LOGIQ E20, Version: R3.1.2
 GE Healthcare LOGIQ Fortis, Version: R3.1.2
 GE Healthcare LOGIQ P5
 GE Healthcare LOGIQ P6⁹
 GE Healthcare LOGIQ P9
 GE Healthcare, Product: LOGIQ P9, Version: v4.5

GE Healthcare LOGIQ S8^{5,9}
GE Healthcare, Product: LOGIQ S8, Version: v4.2
GE Healthcare Venue R2
GE Healthcare VenueGo, v. 302.87.0
GE Healthcare VIVID S6
GE Healthcare VIVID 7⁹
GE Healthcare VOLUSONI (V730)⁹
GE Healthcare VIVID Q⁹
GE Healthcare VIVID IQ
GE Vingmed VIVID E9⁹
GE Healthcare VIVID E90 (v202)²⁰
GE Healthcare VIVID E95 (Version 202)²³
GE Healthcare VIVID S70
GE Healthcare, Device: Vivid e95, Version:204.67.3
GE Healthcare, Product: VScan Air, Version: 1.1
GE Healthcare, Device: Vivid E95, Version:206.x.x
GE Healthcare, Device: Vivid E90, Version:206.x.x
GE Healthcare, Device: Vivid E80, Version:206.x.x
GE Healthcare, Device: Vivid S70N, Version:206.x.x
GE Healthcare, Device: Vivid S60N, Version:206.x.x
GE Healthcare, Device: Vivid iq, Version:206.x.x
GE Healthcare, Device: Vivid T8, Version:206.x.x
GE Healthcare, Device: Vivid T9, Version:206.x.x
GE Healthcare, Device: EchoPAC Software Only, Version:206.x.x
GE Healthcare, Device: EchoPAC Plug-in, Version:206.x.x
GE HealthCare, Product: Venue, Software Version: 304.61
GE HealthCare, Product: Venue Go, Software Version: 304.61
GE HealthCare, Product: Venue Fit, Software Version: 304.61
Hitachi Healthcare Avius^{5,9}
Hitachi Healthcare EUB Hi VISION 5500^{5,9}
Hitachi Healthcare Hi VISION 900^{5,9}
Hitachi Healthcare Hi VISION Preirus
Hitachi Healthcare SSD-ALPHA 10SX
Hitachi Healthcare SSD-ALPHA5
Hitachi Healthcare SSD-ALPHA6
Hitachi Healthcare SSD-ALPHA7
Hitachi Healthcare Arietta 60
Hitachi Healthcare Arietta 65, v.1.0²³
Hitachi Healthcare Arietta 70
Hitachi Healthcare Arietta 850
Hitachi Healthcare ProSound F75
Hitachi Healthcare F37
Hitachi Healthcare NOBLUS
Hitachi Healthcare Precision
Hitachi Healthcare Lisendo 880
Hitachi Healthcare Arietta Prologue
KONICA MINOLTA HS1
Mindray Medical M7^{5,9}
Mindray Medical DC8 Exp^{5,9}
Mindray Medical TE7^{5,9}
Mindray Medical M9^{5,9}
Mindray, Product: TE X, Version: 1.5.1
Mindray, Product: TE7 Series, Version: v5.4.0

Mindray, Product: TE5 Series, Version: v5.4.0
Mindray, Product: TE9 Series, Version: v5.4.0
Mindray, Product: TE10 Series, Version: v5.4.0
Parks Medical Electronics Flo-Lab 2100SX (Sonova SW)⁹
Parks Medical Electronics Flo-Lab 2100SX (SonovaE SW)
Parks Medical Electronics Flo-Lab 2100SX 2 (SonovaE SW)
Philips Affiniti 50G (Version: 2.0.3.396)
Philips iE33₅
Philips iU22_{5,19}
Philips EPIQ5
Philips EPIQ7
Philips Epiq Elite/EPIQ 7G, (Version: 5.0.1.9)³⁰
Philips CX50
Philips Lumify, v. 1.8.3
Philips Lumify, v. 2.0
Philips Sparq 2.0
Philips Sparq 3.5
Philips Healthcare, EPIQ, v9.0.x (software 9.0.1)
Philips Healthcare, EPIQ, v7.0 (software 7.0.8)
Philips, Product: Affiniti, Version: R6.0³⁷
Philips, Product: Affiniti, Version: R9.0³⁷
Philips, Product: IntraSight, Version: 5.2.2.x
Philips, Product: Compact 5000 Series, Version: 1.0
Philips, Product: Affiniti 70, Software Version: R10.0
Philips, Product: Affiniti 50, Software Version: R10.0
Philips, Product: Affiniti 30, Software Version: R10.0
Philips, Product: EPIQ 7, Software Version: R10.0
Philips, Product: EPIQ Elite, Software Version: R10.0
Philips, Product: EPIQ Elite Advanced, Software Version: R10.0
Philips, Product: EPIQ 5, Software Version: R10.0
Philips, Product: EPIQ CVx, Software Version: R10.0
Philips, Product: EPIQ EPIQ CVxi, Software Version: R10.0
Philips, Product: EPIQ Elite, Version: 11.0
Philips, Product: EPIQ Elite Advanced, Version: 11.0
Philips, Product: EPIQ 5, Version: 11.0
Philips, Product: EPIQ 7, Version: 11.0
Philips, Product: EPIQ CVx, Version: 11.0
Philips, Product: EPIQ CVxi, Version: 11.0
Philips, Product: Affiniti 50, Version: 11.0
Philips, Product: Affiniti 70 , Version: 11.0
Philips, Product: Affiniti 30 , Version: 11.0
Philips, Product: Affiniti CVx, Version: 11.0
Samsung RS80
Samsung WS80A
Samsung HM70A
Samsung NeuroLogica, Product: RS85, Version: 2.05.01
Samsung NeuroLogica, Product: HS60, Version: V2.04
Samsung NeuroLogica, Product: V7, Version: v1.03.00
Samsung NeuroLogica, Product: V8, Version: v1.03
Samsung NeuroLogica, Product: HS40, Version: v1.05
Samsung NeuroLogica, Product: HERA W10, Version: v1.03
Samsung NeuroLogica, Product: HM70 EVO, Version: v1.01
Samsung NeuroLogica, Product: HS60, Software Version: v2.05

Samsung NeuroLogica, Product: RS85, Software Version: v2.08
Samsung NeuroLogica, Product: V8, Software Version: v1.04
Samsung NeuroLogica, Product: V7, Software Version: v1.04
Siemens G20^{5,9}
Siemens G50^{5,9}
Siemens G60S^{5,9}
Siemens S1000
Siemens S2000
Siemens S3000
Siemens SC2000
Siemens Sequoia (v12.2)^{5,9}
Siemens (Acuson) X300^{5,9}
Siemens (ACUSON Sequoia) VA10A
Siemens, Product: ACUSON SC2000, Version: VC10
Siemens, Product: ACUSON Sequoia, Version: VA40
Siemens, Product: ACUSON Sequoia, Version: VA30
Sonomed Escalon, Product: VuMAX HD, Version: v4.0.0
Sonomed Escalon, Product: VuPad, Version v4.0.0
SonoSite (FujiFilm) Edge
SonoSite Edge II
SonoSite M-Turbo
SonoSite MicroMAXX
SonoSite (FujiFilm) NanoMaxx
SonoSite Titan
SonoSite S-Series
SonoSite S II
SonoSite X-Porte
SonoSite (FujiFilm) iViz
SonoSite (FujiFilm)-PX, v. 100.80.100.086
SuperSonic Imagine Aixplorer
TeraTech Corp/Terason, uSmart3200T Plus, v. 5.14.4
Toshiba Aplio 80
Toshiba Aplio 300
Toshiba Aplio 400
Toshiba Aplio 500
Toshiba Aplio i700
Toshiba Aplio i800
Toshiba Aplio i900
Toshiba Aplio MX
Toshiba Aplio XG
Toshiba Artida
Toshiba Nemio (SSA-550A)
Toshiba Nemio XG (SSA-580A)
Toshiba Viamo
Toshiba UltraExtend¹
Toshiba Xario
Toshiba Xario XG (SSA-680A)
Toshiba Xario 100/200
Unetixs Vascular MultiLab Series
Ultrasonix Medical Corp. SonicTOUCH
Ultrasonix Medical Corp. Sonix MDP
Ultrasonix Medical Corp. Sonix OP
Ultrasonix Medical Corp. Sonix RP

Ultrasonix Medical Corp. Sonix SP
US-Vascular Vasculab 4040, version 1.0.180
Viasonix Falcon
Viasonix Falcon, v.1.6.0
Volcano S5⁹
Volcano S5i^{5,9}
Zonare Z.One⁹
Zonare Z.OneUltra⁹
Zonare ZS3

Notes

1. No DICOM Modality Worklist capability.
5. Tested as DICOM Object. Test images were not converted to Targa format.
9. Automatic measurements may not be possible; manual calibration may be required.
17. DICOM Query/Retrieve – Study/Root Q/R fully supported.
19. Reported connectivity issues, please contact support for assistance
23. Not approved for DICOM Query/Retrieve.

Visual Field Analyzer (Encapsulated PDF)

Supported VL Visual Field Analyzer (Encapsulated PDF)

Carl Zeiss Meditec, Inc. HFA 720-i
Carl Zeiss Meditec, Inc. HFA 740-i
Carl Zeiss Meditec, Inc. HFA 745-i
Carl Zeiss Meditec, Inc. HFA 750-i
Carl Zeiss Meditec, Inc. HFA II (Humphrey)
Carl Zeiss Meditec, Inc. HFA3 (with Forum)
Carl Zeiss Meditec, Inc. HFA3
Carl Zeiss Meditec, Inc. HFA3V1.3
Carl Zeiss Meditec, Inc. HFA3V1.4
Carl Zeiss Meditec, Inc. HFA3 V1.5
Carl Zeiss Meditec, Inc. HFA3 V1.5.1
Carl Zeiss Meditec, Inc. HFA3 V1.5.2
Carl Zeiss Meditec, Inc. HFA3, Version: 1.6.x
Carl Zeiss Meditec, Inc. Matrix 800 (HFA II)
Carl Zeiss Meditec, Inc. Atlas 9000, V: 3.0.2

VL – Endoscopy

Supported VL – Endoscopy Devices

Arthrex, Device: SynergyUHD4, Version: 4.17.3
Endosoft (UTECH) Endosoft Systems¹⁸
Endosoft (UTECH) Endosoft 6.0 (Platform: 6.0)¹⁸
Endosoft (UTECH) EndoVault 1.3 (Platform: 6.0)¹⁸
Endosoft EndoVault 2.15 and 3.0¹⁸
Envisionier Medical Technologies, Inc. eGoManager
Global Media Group, LLC CapSure 2.0⁵
Infinite Software Solutions Inc D\BVA MD-Reports
Image Stream Medical VaultStream /PatInfoEngine, 10.2.1
Image Stream Medical VaultStream /PatInfoEngine, 10.2.2
Image Stream Medical nCare /PatInfoEngine, 10.2.1
Karl Storz AIDA v1.4
Karl Storz AIDA v1.6
Karl Storz AIDA v1.7
Karl Storz AIDA HD Connect
Karl Storz OR1-AIDA Compact II

Karl Storz StreamConnect NEO¹
KARL STORZ, Product: AIDA Bella, Version: 1.8.x
KayPENTAX KDS (Digital Stroboscopy Workstation)
KayPENTAX 9200D KDS (Digital Stroboscopy Workstation) w/KDCM Module
KayPENTAX 7245D DSW (Digital Swallowing Workstation) w/KDCM Module
KayPENTAX 9310HD Digital Video Capture Module, V. 3.4²⁶
KayPENTAX 9372HD Digital Capture, v1.1
KayPENTAX 9372HD Digital Capture, v1.1.1
MedXchange DRSHD – Digital Recording System
MedXchange DRSHD – Digital Recording System v1.4
MedXchange 4Klarity
MedXchange DRSHD – DRSHD1080P
MedXchange Evolution HD
MedXchange Evolution4K, v. 1.6.4²⁵
MedXchange HDMDPro
Olympus EndoWorks
Olympus America, Inc. Endoworks 7,4
Olympus Knowledge Exchange (KE) System
Olympus, Product: Olympus nCare, Version: 10.4.1.2
STI UltraSight HD
Stryker Endoscopy SDC-HD
Stryker Endoscopy SDC Ultra
Stryker SDC3
Stryker DataMediatOR, V: 1.5.0.0³⁴

Notes

18 DICOM Encapsulated PDF.

25 This device is not approved for the following SOP Classes: Multi-Frame Secondary Capture Storage, Video Endoscopic Image Store and Video Photographic Image Store. Only approved for Secondary Capture Image Storage and VL Endoscopic Image Storage.

VL – Endoscopy/Microscopy

Supported VL – Endoscopy/Microscopy Devices

ImageStream nStream
ImageStream nCare MX
ImageStream nCare RX
ImageStream nStream GX
ImageStream VaultStream

VL – Pathology

Supported VL –Pathology Devices

Apollo Telemedicine ASAP Imaging

XA – X-Ray Angiographic

Supported XA – X-Ray Angiographic devices

GE Healthcare, Device: Allia IGS 5, Software: DL, Version: 21.0.10
GE Healthcare, Device: Allia IGS 7, Software: DL, Version: 21.0.10
GE Healthcare, Device: Allia IGS 7OR, Software: DL, Version: 21.0.10
Philips Medical Systems, Device: Zenition70, Version: 5.3.3
Philips Medical Systems, Device: Zenition50, Version: 5.3.3

Philips Medical Systems, Product: Allura Xper²⁸, Software: Allura Clarity, Version: 8.2

Philips, Product: Azurion²⁸, Version: R3.0

Philips, Product: Zenition 90⁴¹, Version: R7.1

Philips, Product: Zenition 30⁴¹, Version: 6.1

Siemens Healthineers, Product: ARTIS icono floor²⁸, Software Version: ARTIS VE30A

Siemens Healthineers, Product: ARTIS icono biplane²⁸, Software Version: ARTIS VE30A

Siemens Healthineers, Product: ARTIS pheno²⁸, Software Version: ARTIS VE30A

Siemens Healthineers, Product: ARTIS icono floor²⁸, Software Version: ARTIS VE40A

Siemens Healthineers, Product: ARTIS icono biplane²⁸, Software Version: ARTIS VE40A

Siemens Healthineers, Product: ARTIS pheno²⁸, Software Version: ARTIS VE40A

Siemens Healthineers, Product: ARTIS icono ceiling²⁸, Software Version: ARTIS VE40A

XC – External-camera Photography (Encapsulated PDF)

Supported XC – External-camera Photography Devices (Encapsulated PDF)

Global Media Group, LLC CapSure Version 2.1.6.4

GlobalMedia Group, LLC CapSure Version 3.2

GlobalMedia Group, LLC , Device: CapSure, Version: 3.6.7.0

Image Stream Medical nCare /PatInfoEngine, 10.2.1

Notes

The notes below are included for reference purposes; see specific listing for details.

- ¹ No DICOM Modality Worklist capability.
- ² Poor DICOM Modality Worklist capability.
- ³ Tested using the DeJarnette Netshare for interfacing.
- ⁴ Data transferred via an intermediary device. Confirm intermediary device was purchased as well.
- ⁵ Tested as DICOM Object. Test images were not converted to Targa format.
- ⁶ *Obsolete; retained for numbering purposes.*
- ⁷ *Obsolete; retained for numbering purposes.*
- ⁸ Only supplies Modality Worklist.
- ⁹ Automatic measurements may not be possible; manual calibration may be required.
- ¹⁰ Older versions of vendor's software are not supported.
- ¹¹ *Obsolete; retained for numbering purposes.*
- ¹² *Obsolete; retained for numbering purposes.*
- ¹³ *Obsolete; retained for numbering purposes.*
- ¹⁴ *Obsolete; retained for numbering purposes.*
- ¹⁵ Does not support OP SOP Classes.
- ¹⁶ DICOM Query/Retrieve – Performs C-Find SCU and C-Move SCU roles only.
- ¹⁷ DICOM Query/Retrieve – Study/Root Q/R fully supported.
- ¹⁸ DICOM Encapsulated PDF only
- ¹⁹ Reported connectivity issues, please contact support for assistance
- ²⁰ Version of DICOM Software approved on prior
- ²¹ Requires using NNT/iRYS Workstation w/DICOM module
- ²² At the time of validation VA HDIG was out of date with the DICOM Standard and doesn't have the ability to recognize the value, POSTMARKER, the third value of Image Type (0008,0008). Validation testing was done with the Application Entity Security Matrix (AESM) entry configured to: VALIDATE=NO and RELAX VALIDATION=YES.
- ²³ This device is not approved for Query Retrieve
- ²⁴ During testing it was noted that if the objects sent by the vendor have different accession numbers, then the ePDF objects stored in VistA do not display an accession numbers. But if the ePDF objects are made up of images that are from the same accession number, then the ePDF objects stored in VistA display the accession number.
- ²⁵ This device is not approved for the following SOP Classes: Multi-Frame Secondary Capture Storage, Video Endoscopic Image Store and Video Photographic Image Store. Only approved for Secondary Capture Image Storage and VL Endoscopic Image Storage.
- ²⁶ Device does not perform MWL, it is a Vista Storage only device
- ²⁷ Approved HL7 interface 5/2/2014. The latest version 3.5 is not validated for use within VA.
- ²⁸ Since the VistA Imaging Query/Retrieve Provider only supports Study Level at this time and the device cannot perform Study Level retrieves, the device may be used to obtain prior radiology images from a Query/Retrieve provider other than VistA Imaging.
- ²⁹ VistA Imaging does not support biometry related SOP classes such as: Keratometry Measurements Storage (1.2.840.10008.5.1.4.1.1.78.3), Ophthalmic Axial Measurements Storage (1.2.840.10008.5.1.4.1.1.78.7) and Intraocular Lens Calculations Storage (1.2.840.10008.5.1.4.1.1.78.8).
- ³⁰ This device was validated for US SOP Classes only.
- ³¹ Comments: Broad Query matching key attributes is not supported. Can filter by Modality
- ³² VistA Imaging does not support PET (Radiopharmaceutical) Dose SR storage.
- ³³ DCS VG80A_mCT.doc is an addendum to syngoCT-VB10.pdf. Everything valid for syngo CT VB10 is also valid for Biograph mCT. Specifics to mCT are described in VG80A_mCT.doc. The specific SOP Classes tested for the VG80A_mCT are 1.2.840.10008.5.1.4.1.1.2 CT Image Storage, 1.2.840.10008.5.1.4.1.1.128 Positron Emission Tomography Image Storage, 1.2.840.10008.5.1.4.1.1.7 Secondary Capture Image Storage, 1.2.840.10008.5.1.4.1.1.88.67; X-Ray Radiation Dose SR.
- ³⁴ This device has been granted conditional approved for use with VistA Imaging pending the updated Stryker DICOM conformance statement document – to reflect DICOM Transfer Syntax changes as required by the VA and the inclusion of the decision to use Study Instance UID generated by the DataMediator. The Stryker DICOM Conformance Statement will be released in Q4 2021 to include the DICOM Transfer Syntax changes, as required by the VA.
- ³⁵ Validated only Digital Intra-oral X-ray Image Storage - For Presentation SOP class since other SOP classes are out of scope for the purposes of sensor device validation.
- ³⁶ The device is not equipped to create new studies rather sending series to existing studies. So not able to validate MWL. Since work list not created in vender PAC, not able to validate device images through validation process. Performed Q/R and verified SOP class from storage server.
- ³⁷ This device is not approved for Secondary Capture Image Storage.
- ³⁸ VistA Imaging Query/Retrieve Provider only supports Study Level at this time and the device cannot perform Study Level retrieves. Performed additional steps as work-around to achieve the validation with VistA imaging by allowing to include approved modality in the workflow by performing successful worklist query from VistA and acquired image for the Patient and populated the data into workstation. This will allow workstation to create

- secondary capture image after postprocessing the acquired image on workstation and store the images to VistA.
- ³⁹ DICOM Query/Retrieve – Study/Root Q/R not supported.
- ⁴⁰ VISTA imaging doesn't support DOSE SR.
- ⁴¹ For Secondary capture IOD SOP Instance UID Integrity is not applicable as the secondary capture is a screen capture which is generated on the fly and not stored in the system, so every secondary capture is a new image hence new UID is assigned to the object.